

TUV Rheinland of North America Inc.
12 Commerce Rd, Newtown, CT 06470, USA
Product Safety and Quality
Electrical Services

July 17, 2011

Attn: Anderson Micu
Biosignal Group.
760 Parkside Avenue, Room 206, Brooklyn, New York. 11226. USA.

TUV File Number:	31190202.001
Type of Equipment	Patient Worn Electroencephalograph Recorder
Model Designation	microEEG
Test Requirements	IEC 60601-1:1988 + A1:1991 + A2:1995 EN 60601-1:1990 +A1+A2+A13

Dear Micu

We have completed the evaluation of your equipment to the above mentioned requirements and found it to be in compliance.

This test report includes the following documents:

1. IEC/EN 60601-1 Test Report - 46 pages

Attachments to this report:

1. US and Canada National Differences - 15 pages
2. IEC 60601-2-26 Test report, 8 pages
3. IEC 60601-1-1 Test report, 13 Pages
4. Information for users, 86 Pages
5. microEEG with charger EMC test report TUVRNA File #31160175.001, 82 Pages
6. Photos of system –6 pages
7. Schematics – 9 pages

Please contact us for any questions you may have.

Sincerely,
TUV Rheinland of North America, Inc.

William Stearns

